

CE DECLARATION OF CONFORMITY

Plum Eyewash 0.9% Sodium Chloride

Eye irrigation fluid for first aid

We, the manufacturer, of the products covered by this declaration, hereby declare, that the bottles/ampoules with sterile irrigation fluid listed below, meet the requirements in the Council Directive 93/42/EEC of 14 June 1993, the Directive 2007/47/EC and its relevant transposition into national laws of the member states into which we place the device. Plum A/S is responsible for the Declaration of Conformity concerning the medical devices listed below:

Art No	Art Name	Volume
5160	Plum Eyewash 0.9% Sodium Chloride QuickRinse	20 mL
4691	Plum Eyewash 0.9% Sodium Chloride	200 mL
4604	Plum Eyewash 0.9% Sodium Chloride	500 mL
4707	Plum Eyewash 0.9% Sodium Chloride	1000 mL
4861	Plum Eyewash 0.9% Sodium Chloride DUO	500 mL
4800	Plum Eyewash 0.9% Sodium Chloride DUO	1000 mL

With the intended use:

Plum Eyewash 0.9% Sodium Chloride is to be used as a first aid product for eye irrigation.

We guarantee and declare that:

- 1. The device is a Class Is device according to Annex IX, rule 5.
- 2. The irrigation fluid is STERILE
- 3. The planning and manufacture procedures are conducted in conformity with the Company Quality System, in compliance with the provisions of Annex VII combined with Annex V to the Directive.

The notified body Presafe Denmark A/S (ID No. 0543) certifies Plum A/S. The latest EC Certificate No. DGM-272 2019-12-03.

Assens, 2020-04-29

Plum A/S

Lise S Bollen Quality Manager