

134, Wu Kung Road,
WuKu Industrial. Zone,
Taipei County, Taiwan 248
Tel: +886-2-2299-3279
Fax: +886-2-2298-2698

No: EM/2009/90005C

VERIFICATION OF COMPLIANCE

Product Name : Personal Supersonic Stimulator, Ultrasound Therapeutic Massager
Brand Name : N/A
Model No. : KUM-200
Added Model(s) : KUP-200
Applicant : KING ULTRASONIC CO., LTD.
Address of Applicant : No.8, Wu-Kung 5 Rd., Wu-Ku Industrial Park, Taipei, 248 Taiwan, R.O.C.
Based on SGS EMC Test : EM/2009/90005
Report Number(s)
Date of Issue : Feb. 09, 2010
Applicable Standards : EN60601-1-2 : 2007, CISPR 11 : 2003+A1:2004+A2:2006,
EN61000-3-2 : 2006, EN61000-3-3 : 2008,
IEC61000-4-2 : 1995+A1:1998+A2: 2000,
IEC61000-4-3 : 2006+A1:2007, IEC61000-4-4 : 2004,
IEC61000-4-5 : 2005, IEC61000-4-6 : 2008,
IEC61000-4-8 : 1993+A1:2000, IEC61000-4-11 : 2004

Conclusion

Based upon a review of the Technical Construction File, the apparatus is in compliance with below requirements of:

Medical Devices Directive 93/42/EEC amended by 2007/47/EC

Authorized Signatory:



SGS TAIWAN LTD.
Victor Wen

