

“EC” DECLARATION OF CONFORMITY

Whirl Wheelchairs inc Spares & Accessories


91440320 WHIRL38TR Whirl Attendant Propelled Wheelchair Seat Width 38 cm
91440346 WHIRL41TR Whirl Attendant Propelled Wheelchair Seat Width 41 cm
91440361 WHIRL43TR Whirl Attendant Propelled Wheelchair Seat Width 43 cm
91440379 WHIRL45TR Whirl Attendant Propelled Wheelchair Seat Width 45 cm
91440395 WHIRL48TR Whirl Attendant Propelled Wheelchair Seat Width 48 cm
91440262 WHIRL38SP Whirl Self Propelled Wheelchair Seat Width 38 cm
91440270 WHIRL41SP Whirl Self Propelled Wheelchair Seat Width 41 cm
91440288 WHIRL43SP Whirl Self Propelled Wheelchair Seat Width 43 cm
91440296 WHIRL45SP Whirl Self Propelled Wheelchair Seat Width 45 cm
91440304 WHIRL48SP Whirl Self Propelled Wheelchair Seat Width 48 cm

complies with the assessment criteria of Council's Directives:

- 93/42/EEC, as amended by all subsequent Directives including 2007/47/EC.
- EN12182 : 1999 - Technical aids for disabled persons
- ISO 10993 Biological evaluation of medical devices
- BSEN12183 – Manual Wheelchairs. Requirements & Test Methods
- EN 14971 Risk Management
- ISO 8191-1 Flammability – Cigarette
- ISO 8191-2 Flammability – Match
- and the quality system of the manufacturer is in conformity with the following standard :
 - **BS EN ISO 9001:2008**
- Patterson Medical (Huthwaite) Certificate N° FM36058 originally registered by BSI on 24/01/1997
 - Transferred to LRQA Certificate N° LRQ 4006200 on 11/09/13
 - **BS EN ISO 13485 : 2003**
- Patterson Medical (Bridgend) N° MD76176 originally registered by BSI on 24/12/2003
- - Certificate for approval of full Quality Assurance System

Additional information : Product file No 263B Whirl Wheelchairs
Directive classification : Class I, Non-sterile
Intended Use : For use by persons with restricted movement

Patterson Medical,
Nunn Brook Road, Huthwaite,
Sutton-in-Ashfield, Notts.
NG17 7HU
9th November 2014



P Dudley
Regulatory and Compliance
Manager