



1<sup>st</sup> Floor & South Wing, 2<sup>nd</sup> Floor of Block A and North Wing, 1<sup>st</sup> Floor of Block E, Hourui Third Industrial Park, Xixiang, Bao'an District, Shenzhen City, Guangdong Province 518102, P.R. China

**limoss (Shenzhen) Co., Ltd.**

Tel: +86 0755-29798877  
www.limoss.de  
limoss@limoss.de

## EC Declaration of Conformity

As the solely responsible entity, we hereby declare that the products as listed below meet the basic requirements of the Medical Device Directive **93/42/EEC**, Low Voltage Directive **2014/35/EU**, the Electromagnetic Compatibility Directive **2014/30/EU** and the ROHS II Directive **2011/65/EU**.

Any modification or change of the product without our approval will void the validity of this declaration.

**We are not a manufacturer of medical devices but we have our products tested for the use in medical devices.**

**Valid for the following products from date of construction:** September 9<sup>th</sup> 2016

**Product: Hand Control**

**Model name: HC100/HC110/HC112/HC160**

**The identification of HC100/HC110/HC112/HC160:**



### Applied harmonized standards:

EN 60335-1:2012+A11:2014  
EN 55014-1:2006+A1+A2  
EN 55014-2:2015  
EN 50581:2012

EN 60950-1:2006+A11+A1+A12+A2  
EN 61000-3-2:2014  
EN 61000-3-3:2013  
EN 62233:2008

Shenzhen, September 9<sup>th</sup> 2016

Color Jiang (QRC Department Manager)